



Eckert & Ziegler

Isotope Products

Medical & Isotrak Laboratory  
24937 Avenue Tibbitts • Valencia, CA 91355

Industrial Laboratory  
1800 North Keystone Street • Burbank, CA 91504  
www.ezag.com  
Tel 661.309.1010 • Fax 661.257.8303

## NOMINAL SOURCE CERTIFICATE

**Product Code:** PF24RW-057-20M **Quantity:** 1  
**Model:** PF24RW **SS&DR No.:** CA0406S180S  
**Active Region:** 610 mm x 420 mm (24.0" x 16.5") **ISO/ANSI Classification:** ISO/99/C22414  
**Capsule Type:** EZIP drawing #3515 **Nuclide Half Life:** 271.79 ± 0.09 days  
**Cover/Backing:** Vinyl backed nylon **Recommended Replacement Date:** 2021-09-01

Nuclide	Serial Number	Activity	Reference Date
Co-57	2131-112	740 MBq (20 mCi)	2020-03-01

### Source Field Uniformity Measurement:

Source uniformity measurement of the 122/136 keV gamma emission was performed using a gamma camera scanning system. An array of "unit cells" was measured to calculate the uniformity data shown below.

Unit Cell Area: 0.91 cm<sup>2</sup> Integral Non-Uniformity (INU): 2.00%  
 Coefficient of Variation: 0.60% Differential Non-Uniformity (DNU): 1.65%

**Impurities:** Co-56 and Co-58 combined is < 0.12% on 2020-03-01.

### Notes:

This document uses the numerical convention where 1.000 = 1 and 1,000 = 10<sup>3</sup>.

This document uses the date convention YYYY-MM-DD in accordance with ISO 8601.

Nuclear data was taken from IAEA-TECDOC-619, 1991.

The referenced ISO/ANSI classification is compliant to ISO 2919:2012.

U.S. Patent #7,233,012.

Coefficient of Variation: The standard deviation of The distribution of The measured values by The mean of measured values.

Integral Non-Uniformity: (Max - Min)/(Max + Min), where Max represents the largest measurement and Min represents the smallest measurement in the useful region.

Differential Non-Uniformity: (Max - Min)/(Max + Min) represents the largest deviation between a central value and the eight surrounding measurements.

Leak Test Frequency = 6 months.

Nature of Active Deposit: Co-57 dispersed in resin matrix.

### Leak Test:

The leak test(s) stated below was (were) either derived or directly taken from the leak test methods listed in ISO 9978:1992. The leak test(s) below complies (comply) with ISO 9978:1992 and does (do) not exceed the regulatory limit of <5 nCi (185 Bq) of removable alpha and beta-gamma emitting activities. Leak tests conducted resulted in <5nCi (185 Bq) of removable activity unless otherwise stated on this certificate.

This source was wiped over its entire surface with a moistened filter paper disk. After drying, the disk was checked for activity using a scintillation detector.

SIGNATURE

2020-01-09

DATE

NOTEBOOK: 2130-015

ISO 13485 CERTIFIED



### Authorized Representative

Eckert & Ziegler Nuclitec GmbH Gieselweg 1 38110 Braunschweig GERMANY  
 Tel: +49 (0) 5307 9320 • Fax: +49 (0) 5307 932 293



IPLE 00000001.000