



BS EN 1789:2020 Test Summary - Component

This is to certify that the following ambulance component, tested on a platform used to represent mounting within a vehicle, complies with BS EN 1789:2020, paragraphs 4.4.11 and 5.3 dynamic tests. Component suitable for a Type A/B/C ambulance.

1. Trade name or mark of the device: Stryker Medical
2. Manufacturer's name for the type of device: 6100 M-1 cot combined with:

6381 OR 6373 OR 6376 fasteners, AND:
6060-160-010 EMS Restraint Package, OR:
6500002030 G Rated Restraint Package
3. Manufacturer's name and address:

Stryker Medical
3800 E. Centre Ave.
Portage, MI 49002
United States of America
4. Submitted for test on: As before and 20 September 2021
5. Number of test report: As before [T1789 USQ316199] and T1789-USY536037
6. Date of test report: As before [23 March 2015] and 21 September 2021
7. Authorised designated official: CHRIS McCABE


8. Signed:
9. Date: 20 DECEMBER 2021

Remarks: This test summary has been issued in part based on testing performed to BS EN 1789: 2007 + A2-2014. As the 2020 update made no changes to test methods or requirements the previous testing cited also validates compliance to BS EN 1789:2020.