

Certificate of Analysis

Material : 83625.320
Batch : 14J030523

Toluene HiPerSolv CHROMANORM for HPLC
Expires end of 10/2019

<i>CHARACTERISTICS</i>	<i>SPECIFICATIONS</i>	<i>MEASURED VALUES</i>
Assay (GC)	Min. 99,8 %	100,0 %
Acidity	Max. 0,0005 meq/g	Max. 0,0005 meq/g
Alkalinity	Max. 0,0002 meq/g	Max. 0,0002 meq/g
Evaporation residue	Max. 5 ppm	< 1 ppm
Water	Max. 0,05 %	0,03 %
Transmittance (300 nm)	Min. 70 %	81 %
Transmittance (310 nm)	Min. 80 %	91 %
Transmittance (330 nm)	Min. 95 %	97 %
Transmittance (350 nm)	Min. 98 %	100 %

We certify that this batch conforms to the specifications listed above.

BDL : Below detected limit.

Isabelle Habay, Head of laboratory - Briare
VWR International
Document printed on 03/2015

This document has been produced electronically and is valid without a signature.